



BODYSTEEL & SILVER CO.,LTD

63 Moo 5, Soi Wat Bang-Co Yaek Rong-Foam, Kanchanaphisek Road
Bang Ma-Nang, Bang Yai, NONTABURI 11140 THAILAND

☎ 66 2-1505999 📠 66-2-1505990

sales@steelsilver.com

EO STERILIZATION Quality Control Report

Date : 14-NOV-2023

LOT : 908

STERILIZATION DATE : 7-Nov-2023

EXPIRE DATE : 7-Nov-2028

QC REPORT :

PASS as the test result has been shown as pictures below.

BIOLOGICAL INDICATORS :

The EZTest indicator is placed in a suitable test pack which is representative of the load. The load is then processed as usual.

After sterilization the chemical indicator ring on the label changes from blue to brownish when exposed to Ethylene Oxide (EO gas). This distinguishes exposed from unexposed units (it does not indicate acceptable sterilization).

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.

Result
PASS




Sample Fail



Yellow

Sample New
Indicator



Pink

Quality Assurance By :



Date : 14-NOV-2023

LOT NO 908		Welcome use bangkang medical endo	
Ethylene oxide sterilize		Ster T: 50 °C	
Ster H: 55 x		Ster time: 6h 00min	
Clean time: 5 min		Clean P: -50 kPa	
Clean time: 15 min			
T	P	H	% time
Start: 24.1	-0.3	86.3	12:10
Water: 24.2	-0.2	74.9	12:11
Vacu: 24.2	-5.0	74.4	12:11
Preheat: 35.0	-3.0	74.7	12:25
Preheat: 35.5	-2.0	74.9	12:35
Vacu: 34.4	-60.0	72.3	12:40
Keep: 40.1	-38.8	74.3	12:48
Inta: 39.6	-4.9	69.2	12:51
Vacu: 39.5	-60.0	68.7	12:56
Humid: 39.5	-60.0	68.7	12:56
Humid: 39.3	-57.8	74.8	13:26
Dosi: 37.6	-55.7	70.8	13:35
Rst: 37.8	-50.7	73.1	13:35
Diffu: 38.8	-24.3	74.0	13:45
Heat: 50.1	-14.3	74.8	14:21
Steri: 50.1	-14.3	74.8	14:21
Steri: 52.5	-11.6	74.8	14:31
Steri: 51.4	-8.9	74.8	14:41
Steri: 50.7	-7.4	74.4	14:51
Steri: 50.1	-6.9	74.4	15:11
Steri: 50.4	-6.4	74.8	15:21
Steri: 50.1	-5.9	74.0	15:31
Steri: 50.1	-5.4	74.6	15:41
Steri: 50.1	-4.9	74.6	15:51
Steri: 50.3	-3.7	74.6	16:01
Steri: 50.2	-3.0	74.0	16:21
Steri: 50.4	-2.4	74.2	16:31
Steri: 50.3	-1.9	74.8	16:41
Steri: 50.2	-1.5	74.6	16:51
Steri: 50.1	-1.2	74.4	17:01
Steri: 50.4	-0.9	74.8	17:11
Steri: 50.2	-0.7	74.4	17:21
Steri: 50.2	-0.6	74.4	17:31
Steri: 50.2	-0.5	74.0	17:41
Steri: 50.3	-0.4	74.4	17:51
Steri: 50.2	-0.4	74.2	18:01
Steri: 50.1	-0.4	74.4	18:11
Steri: 50.2	-0.4	74.6	18:21
Steri: 50.1	-0.4	74.6	18:31
Steri: 49.9	-0.5	74.2	18:41
Steri: 50.2	-0.4	74.6	18:51
Steri: 50.3	-0.5	74.2	19:01
Steri: 50.1	-0.4	74.2	19:11
Steri: 50.0	-0.4	74.2	19:21
Steri: 50.1	-0.4	74.2	19:31
Steri: 50.3	-0.4	74.4	19:41
Steri: 50.1	-0.3	74.4	19:51
Steri: 50.1	-0.4	74.8	20:11
Steri: 50.1	-0.5	74.6	20:21
Clean Process			
1th : 51.0	16.8	59.5	20:54
2th : 50.9	16.5	55.2	21:25
3th : 50.7	16.6	54.7	21:57
4th : 50.6	16.5	53.8	22:28
5th : 50.7	16.5	53.1	23:00
Inta: 50.8	-3.7	51.5	23:00
End: 50.8	-9.7	51.5	23:00
Sterilization Completed			
Steril Time: 06h45min			
Usecount: 00098			
Date: 2023-11-7			
Number: R0000450051			
Operator:			

Protocol :

