



BODYSTEEL & SILVER CO.,LTD
63 Moo 5, Soi Wat Bang-Co Yaek Rong-Foam, Kanchanaphisek Road
Bang Ma-Nang, Bang Yai, NONTABURI 11140 THAILAND

☎ 66 2-1505999 📠 66-2-1505990

sales@steelsilver.com

EO STERILIZATION Quality Control Report

Date : 16-Nov-2023

LOT : 0910

STERILIZATION DATE : 09-Nov-2023

EXPIRE DATE : 09-Nov-2028

QC REPORT :

PASS as the test result has been shown as pictures below.

BIOLOGICAL INDICATORS :

The EZTest indicator is placed in a suitable test pack which is representative of the load. The load is then processed as usual.

After sterilization the chemical indicator ring on the label changes from blue to brownish when exposed to Ethylene Oxide (EO gas). This distinguishes exposed from unexposed units (it does not indicate acceptable sterilization).

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.

Result
PASS



Sample Fail




Yellow

Sample New Indicator



Pink

Quality Assurance By:



Date : **16-NOV-2023**

STERI :45-C BAT :0910		09/11/2023 16:13		STEP		TIME		TEMP (C)		PRES (KPa)		HUMI (GRH)	
PH		16:13		28.4		00		92.7					
PV		17:56		45.0		00		93.0					
HU		18:09		45.0		-85		81.9					
GS		18:19		44.4		-84		86.3					
ST		18:21		45.1		-51		NONE					
ST		18:51		45.2		-50		NONE					
ST		19:21		44.6		-50		NONE					
ST		19:51		45.1		-50		NONE					
ST		20:21		45.2		-50		NONE					
ST		20:51		44.6		-50		NONE					
ST		21:21		44.9		-50		NONE					
ST		21:51		45.5		-49		NONE					
ST		22:21		44.8		-49		NONE					
ST		22:51		44.6		-49		NONE					
AE		23:22		45.2		-88		NONE					
AE		00:19		44.8		-88		NONE					
AE		01:18		45.6		-88		NONE					
AE		02:17		44.8		-88		NONE					
AE		03:15		45.6		-87		NONE					
AE		04:14		44.8		-88		NONE					
AE		05:12		45.6		-88		NONE					
AE		06:10		44.9		-88		NONE					
AE		07:08		45.5		-88		NONE					
AE		08:05		44.9		-88		NONE					
AE		09:02		45.5		-88		NONE					
FI		09:22		44.8		-75		NONE					
CHK BY :													
APP BY :													
RESULT :													

Protocol: