



BODYSTEEL & SILVER CO.,LTD
63 Moo 5, Soi Wat Bang-Co Yaek Rong-Foam, Kanchanaphisek Road
Bang Ma-Nang, Bang Yai, NONTHABURI 11140 THAILAND

☎ 66 2-1505999 📠 66-2-1505990

sales@steelsilver.com

EO STERILIZATION Quality Control Report

Date : **25-JAN-2024**

LOT : 920

STERILIZATION DATE : 17-JAN-2024

EXPIRE DATE : 17-JAN-2028

QC REPORT :

PASS as the test result has been shown as pictures below.

BIOLOGICAL INDICATORS :

The EZTest indicator is placed in a suitable test pack which is representative of the load. The load is then processed as usual.

After sterilization the chemical indicator ring on the label changes from blue to brownish when exposed to Ethylene Oxide (EO gas). This distinguishes exposed from unexposed units (it does not indicate acceptable sterilization).

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.

Result
PASS



Sample Fail



Yellow


Sample New Indicator



Pink

Quality Assurance By:

Date: 25-JAN-2024



-STERILIZATION- 17/01/2024 16:04		LOT NO. 920			
STEP	TIME	TEMP(C)	PRES(KPa)	HUMIDITY	Q200
PH	16:04	31.0	00	86.1	
PV	17:20	45.0	00	85.1	
HU	17:42	45.0	-85	76.6	
GS	17:52	44.6	-84	79.8	
ST	17:54	45.3	-51	NONE	
ST	18:24	45.3	-50	NONE	
ST	18:54	44.5	-50	NONE	
ST	19:24	45.1	-50	NONE	
ST	19:54	45.3	-49	NONE	
ST	20:24	44.6	-49	NONE	
ST	20:54	44.9	-49	NONE	
ST	21:24	45.5	-49	NONE	
ST	21:54	45.2	-49	NONE	
ST	22:24	44.6	-49	NONE	
AE	22:54	45.1	-49	NONE	
AE	23:53	45.5	-88	NONE	
AE	00:53	45.8	-88	NONE	
AE	01:53	44.8	-88	NONE	
AE	02:52	45.5	-88	NONE	
AE	03:52	45.2	-88	NONE	
AE	04:51	45.2	-88	NONE	
AE	05:50	45.2	-88	NONE	
AE	06:49	45.2	-88	NONE	
AE	07:48	45.2	-88	NONE	
AE	08:46	45.2	-88	NONE	
FT	08:54	45.5	-76	NONE	

Protocol: