

EO STERILIZATION Quality Control Report

Date : 13-Sep-2024

LOT : 0986

STERILIZATION DATE : 7-Sep-2024

EXPIRE DATE : 7-Sep-2029

QC REPORT :

PASS as the test result has been shown as pictures below.

BIOLOGICAL INDICATORS :

The EZTest indicator is placed in a suitable test pack which is representative of the load. The load is then processed as usual.

After sterilization the chemical indicator ring on the label changes from blue to brownish when exposed to Ethylene Oxide (EO gas). This distinguishes exposed from unexposed units (it does not indicate acceptable sterilization).

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.

Result
PASS



Sample Fail



Yellow

Sample New
Indicator



Pink

Quality Assurance By:

Date: 13-Sep-2024



Lot 986

We have used sea king medical equ
Ethylene oxide sterilization
Start T: 50 °C
Stop T: 55 °C
Start time: 8 hr (min)
Exposure: 8 hr
Clean P: 0 kPa
Exposure: 15 min

T °C	P kPa	H %	time
Start:	24.1	-0.3	96.9 14:53
Water:	24.1	-0.3	74.7 14:53
Phet:	24.1	-0.2	74.1 14:53
Phet:	35.0	-0.2	72.0 15:09
Khat:	35.4	-1.1	74.0 15:19
Vacu:	33.9	-0.1	74.0 15:24
Temp:	40.3	-0.6	74.2 15:33
Intra:	39.7	-0.4	74.2 15:36
Vacu:	38.7	-0.1	74.0 15:41
Temp:	38.7	-0.1	74.0 15:41
Khat:	38.2	-0.5	74.2 15:41
Cost:	37.5	-0.7	74.0 16:19
Stop:	37.7	-0.1	74.2 16:19
Diffu:	37.8	-0.2	74.0 16:29
Khat:	50.2	-0.3	74.7 17:04
Start:	50.0	-0.3	74.9 17:04
Start:	52.0	-0.6	74.1 17:14
Start:	50.9	-0.3	74.7 17:24
Start:	50.7	-0.7	74.3 17:34
Start:	50.4	-0.2	74.9 17:44
Start:	50.2	-0.1	74.9 17:54
Start:	50.1	-0.0	74.9 18:04
Start:	50.2	-0.0	74.9 18:14
Start:	50.2	-0.0	74.1 18:24
Start:	50.1	-0.0	74.1 18:34
Start:	50.3	-0.0	74.5 18:44
Start:	50.2	-0.0	74.5 18:54
Start:	50.0	-0.1	74.1 19:04
Start:	50.4	-0.1	74.3 19:14
Start:	50.1	-0.1	74.3 19:24
Start:	50.4	-0.1	74.5 19:34
Start:	50.4	-0.2	74.1 19:44
Start:	50.2	-0.2	74.9 19:54
Start:	50.3	-0.3	74.7 20:04
Start:	50.1	-0.3	74.5 20:14
Start:	50.1	-0.3	74.5 20:24
Start:	50.4	-0.3	74.6 20:34
Start:	50.2	-0.4	74.5 20:44
Start:	50.3	-0.4	74.5 20:54
Start:	50.1	-0.5	74.3 21:04
Start:	50.1	-0.5	74.9 21:14
Start:	50.2	-0.5	74.9 21:24
Start:	50.1	-0.6	74.1 21:34
Start:	50.1	-0.6	74.5 21:44
Start:	50.2	-0.6	74.5 21:54
Start:	50.2	-0.6	74.5 22:04
Start:	50.0	-0.6	74.3 22:14
Start:	50.4	-0.5	74.7 22:24
Start:	50.2	-0.5	74.7 22:34
Start:	50.2	-0.5	74.9 22:44
Start:	50.3	-0.5	74.9 22:54

Clean Process
1th: 50.4 -17.5 74.3 23:30
2th: 50.6 -17.3 74.0 23:57
3th: 50.4 -17.2 74.9 01:24
4th: 50.4 -16.7 74.1 01:50
5th: 50.4 -16.9 74.3 01:17
6th: 50.4 -17.2 74.2 01:44
7th: 50.4 -17.2 74.4 02:11
8th: 50.4 -17.2 74.4 02:39
9th: 50.4 -10.6 74.9 02:38
End: 50.4 -0.6 74.9 02:38

Sterilization Completed
Start Time: 11h45min
Date: 2012-9-7
Batch: 0024-97
Operator: 5010000000001

Protocol: