



Date : 24-Apr-2026

EXPIRE DATE : 22-Apr-2030

PASS as the test result has been shown as pictures below.

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.




Yellow



Pink

Date: 24-Apr-2026



Protocol: