



Date : 06-Jun-2026

EXPIRE DATE : 30-May-2031

PASS as the test result has been shown as pictures below.

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.



Protocol: