



Date : 31-July-2026

EXPIRE DATE : 23-July-2031

PASS as the test result has been shown as pictures below.

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.



Yellow

Pink

Date: 31-July-2026

W. Tamee

[illegible]

Protocol: