



Date : 7-Aug-2026

EXPIRE DATE : 31-July-2031

PASS as the test result has been shown as pictures below.

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.

Pink

Quality Assurance By:

W. Tamee

STEEL AND SILVER CO. LTD. BODY

Protocol: