



Date : 7-Aug-2026

EXPIRE DATE : 4-Aug-2031

PASS as the test result has been shown as pictures below.

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.



Yellow



Pink

Date: 7-Aug-2026

W. Tamee

[illegible]

Protocol: