



Date : 31-Aug-2026

EXPIRE DATE : 27-Aug-2031

PASS as the test result has been shown as pictures below.

After the incubation period of 48 hours, the appearance of a yellow color indicates bacterial growth. No color change indicates adequate sterilization.



Date: 31-Aug-2026

Protocol: